


SECTION B-B

NOTE:  
1. LEAK RATE: <math> < 1 \times 10^{-9}</math> ATM-CC/SEC HE

ITEM NO.	DESCRIPTION	QTY
1	FLANGE, 304 STN STL	1
2	14KV SUB ASSEMBLY	4
3	CONDUCTOR, NICKEL 200	4

Allectra part No. 264-24NI15K-4-K50

UNLESS OTHERWISE SPECIFIED:	QUOTATION #:		APPROVED
	DIMENSIONS ARE IN INCHES [ ] DIMENSIONS ARE IN MILLIMETERS FOR REFERENCE  TOLERANCES: ANGULAR: MACH ± 1/2° TWO PLACE DECIMAL ± .030 [.76] THREE PLACE DECIMAL ± .015 [.38]  REMOVE ALL BURRS AND SHARP EDGES MIN SURFACE FINISH: Ra 63 µin   Ra 1.6 µm   N7 THIRD ANGLE PROJECTION: 	DRAWN:	NAME: AP
CHECKED:		NAME: JB	DATE: 7/20/2021
APPROVED:		NAME: AP	DATE: 7/20/2021
APPLICATION			
USED ON:		BASED ON:	
MOUNT: Ø2.95" QF (QF 50)	WEIGHT: 0.446	TITLE: 4X 14KV POWER FEEDTHROUGH 2.16" QUICK FLANGE MOUNT  DWG NO: A0506-1-SP-2.95QF REV:	
Tuesday, July 20, 2021 1:36:19 PM - BY: apettay			